

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110818\
 Data File : VX005850.D
 Acq On : 08 Nov 2018 16:49
 Operator : JC/MD
 Sample : J5858-11
 Misc : 6.67µ/5mL/100µl/5.00mL/MSVOA_X/MEOH
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SB-9-S

Quant Time: Nov 09 09:31:48 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	114390	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	179464	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	158603	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	76818	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	69813	53.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.18%	
35) Dibromofluoromethane	5.50	113	56392	46.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.40%	
50) Toluene-d8	8.72	98	202199	49.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.66%	
62) 4-Bromofluorobenzene	11.14	95	75772	51.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.86%	
Target Compounds						
39) Methylcyclohexane	7.45	83	22575	12.874	ug/l #	90
73) Isopropylbenzene	11.02	105	8916	2.007	ug/l	98
78) n-propylbenzene	11.36	91	9584	1.831	ug/l	93
85) sec-Butylbenzene	11.94	105	29277	6.269	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110818\
Data File : VX005850.D
Acq On : 08 Nov 2018 16:49
Operator : JC/MD
Sample : J5858-11
Misc : 6.67µ/5mL/100ul/5.00mL/MSVOA_X/MEOH
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SB-9-S

Quant Time: Nov 09 09:31:48 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration

