


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX110818\
Data File : VX005838.D
Acq On : 08 Nov 2018 09:46
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Title : SW846 8260
Last Update : Wed Nov 07 14:40:53 2018



AutoFind: Scans 1636, 1637, 1638; Background Corrected with Scan 1630

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.1	2791	PASS
75	95	30	60	51.2	7132	PASS
95	95	100	100	100.0	13917	PASS
96	95	5	9	6.8	950	PASS
173	174	0.00	2	1.6	200	PASS
174	95	50	100	89.2	12410	PASS
175	174	5	9	7.6	942	PASS
176	174	95	101	96.7	12006	PASS
177	176	5	9	6.5	777	PASS