

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110819\
 Data File : VX013356.D
 Acq On : 08 Nov 2019 14:42
 Operator : JC/SP
 Sample : K5765-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT-TB-20191107

Quant Time: Nov 11 03:29:26 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 02 07:06:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	848668	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1325846	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	1186207	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	549796	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	483810	48.12	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.24%	
35) Dibromofluoromethane	5.48	113	403099	48.47	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.94%	
50) Toluene-d8	8.70	98	1550244	49.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.90%	
62) 4-Bromofluorobenzene	11.13	95	535637	47.34	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.68%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110819\
Data File : VX013356.D
Acq On : 08 Nov 2019 14:42
Operator : JC/SP
Sample : K5765-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT-TB-20191107

Quant Time: Nov 11 03:29:26 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 02 07:06:38 2019
Response via : Initial Calibration

