

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110819\
 Data File : VX013357.D
 Acq On : 08 Nov 2019 15:05
 Operator : JC/SP
 Sample : K5765-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT-011-IDWGW-20191107

Quant Time: Nov 11 03:32:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 02 07:06:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	817857	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1266583	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	1136750	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	533746	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	458144	47.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.56%	
35) Dibromofluoromethane	5.48	113	378373	47.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.26%	
50) Toluene-d8	8.71	98	1509459	50.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.80%	
62) 4-Bromofluorobenzene	11.13	95	509751	47.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.32%	
Target Compounds						
16) Acetone	2.43	43	11422	2.363	ug/l	97
60) Dibromochloromethane	9.57	129	4662	0.524	ug/l	99
64) Tetrachloroethene	9.33	164	2285	0.281	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110819\
Data File : VX013357.D
Acq On : 08 Nov 2019 15:05
Operator : JC/SP
Sample : K5765-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT-011-IDWGW-20191107

Quant Time: Nov 11 03:32:06 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 02 07:06:38 2019
Response via : Initial Calibration

