

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110819\
 Data File : VX013358.D
 Acq On : 08 Nov 2019 15:29
 Operator : JC/SP
 Sample : K5765-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT-010-IDWGW-20191107

Quant Time: Nov 11 03:37:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 02 07:06:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	815224	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1265935	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	1113399	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	516524	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	459414	47.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.14%	
35) Dibromofluoromethane	5.48	113	379664	47.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.64%	
50) Toluene-d8	8.71	98	1494331	49.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.84%	
62) 4-Bromofluorobenzene	11.13	95	506448	46.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.76%	
Target Compounds						
16) Acetone	2.43	43	6789	1.409	ug/l #	73
60) Dibromochloromethane	9.58	129	4405	0.496	ug/l	73
64) Tetrachloroethene	9.33	164	2959	0.371	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110819\
Data File : VX013358.D
Acq On : 08 Nov 2019 15:29
Operator : JC/SP
Sample : K5765-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT-010-IDWGW-20191107

Quant Time: Nov 11 03:37:54 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 02 07:06:38 2019
Response via : Initial Calibration

