

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110821\
 Data File : VX025094.D
 Acq On : 08 Nov 2021 13:55
 Operator : JC/MD
 Sample : VX1108MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VBLK631

Quant Time: Nov 09 04:13:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML110821WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Nov 09 03:59:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	189408	50.000	ug/L	# 0.00
28) Chlorobenzene-d5	10.055	117	204754	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	73781	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	86452	50.477	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	= 100.960%	
7) Chloroethane-d5	1.666	69	61834	57.589	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	= 115.180%	
11) 1,1-Dichloroethene-d2	2.307	63	135696	41.134	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 82.260%	
21) 2-Butanone-d5	4.453	46	130289	110.690	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	= 110.690%	
24) Chloroform-d	5.056	84	175558	52.140	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 104.280%	
26) 1,2-Dichloroethane-d4	5.958	65	99523	45.183	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 90.360%	
32) Benzene-d6	5.977	84	233379	32.012	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 64.020%#	
36) 1,2-Dichloropropane-d6	7.306	67	88263	41.087	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	= 82.180%	
41) Toluene-d8	8.653	98	234555	38.274	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 76.540%#	
43) trans-1,3-Dichloroprop...	8.952	79	47239	41.201	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 82.400%	
47) 2-Hexanone-d5	9.385	63	95181	103.728	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	= 103.730%	
56) 1,1,2,2-Tetrachloroeth...	11.195	84	113112	38.039	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	= 76.080%	
66) 1,2-Dichlorobenzene-d4	12.323	152	76921	53.190	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 106.380%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110821\
Data File : VX025094.D
Acq On : 08 Nov 2021 13:55
Operator : JC/MD
Sample : VX1108MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VBLK631

Quant Time: Nov 09 04:13:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML110821WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Nov 09 03:59:51 2021
Response via : Initial Calibration

