

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110821\
 Data File : VX025103.D
 Acq On : 08 Nov 2021 18:17
 Operator : JC/MD
 Sample : M4464-06ME 10X
 Misc : 6.74g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GB7K4ME

Quant Time: Nov 09 04:16:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML110821WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Nov 09 03:59:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.769	114	194984	50.000	ug/L	# 0.00
28) Chlorobenzene-d5	10.061	117	184209	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	94005	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.367	65	80434	45.621	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	91.240%	
7) Chloroethane-d5	1.660	69	36936	33.417	ug/L	-0.01
Spiked Amount	50.000	Range 70 - 130	Recovery	=	66.840%#	
11) 1,1-Dichloroethene-d2	2.306	63	128208	37.752	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	75.500%	
21) 2-Butanone-d5	4.464	46	122888	101.417	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	101.420%	
24) Chloroform-d	5.062	84	147265	42.486	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	84.980%	
26) 1,2-Dichloroethane-d4	5.958	65	116130	51.214	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	102.420%	
32) Benzene-d6	5.976	84	265367	40.460	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	80.920%	
36) 1,2-Dichloropropane-d6	7.311	67	99599	51.535	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	103.080%	
41) Toluene-d8	8.653	98	277309	50.297	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	100.600%	
43) trans-1,3-Dichloroprop...	8.951	79	50862	49.309	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	98.620%	
47) 2-Hexanone-d5	9.390	63	71414	86.507	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	86.510%	
56) 1,1,2,2-Tetrachloroeth...	11.195	84	126124	47.146	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	94.300%	
66) 1,2-Dichlorobenzene-d4	12.323	152	88136	47.833	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	95.660%	
Target Compounds						Qvalue
10) 1,1,2-Trichloro-1,2,2-...	2.331	101	6176	3.902	ug/L #	85
12) 1,1-Dichloroethene	2.318	96	4396	3.158	ug/L #	1
13) Acetone	2.392	43	9980	8.370	ug/L	96
19) 1,1-Dichloroethane	3.617	63	20311	6.884	ug/L	98
20) cis-1,2-Dichloroethene	4.495	96	140753	86.009	ug/L	73
30) 1,1,1-Trichloroethane	5.391	97	35506	10.640	ug/L #	91
34) Trichloroethene	7.129	95	72299	40.949	ug/L	95
42) Toluene	8.720	91	1119452	169.647	ug/L	100
53) m,p-Xylene	10.305	106	4004	1.552	ug/L	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110821\
Data File : VX025103.D
Acq On : 08 Nov 2021 18:17
Operator : JC/MD
Sample : M4464-06ME 10X
Misc : 6.74g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GB7K4ME

Quant Time: Nov 09 04:16:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML110821WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Nov 09 03:59:51 2021
Response via : Initial Calibration

