

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110822\
 Data File : VX032644.D
 Acq On : 08 Nov 2022 11:21
 Operator : JC/MD
 Sample : VX1108WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1108WBL01

Quant Time: Nov 09 04:35:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	117556	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	193218	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	172256	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83320	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	81162	49.689	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.380%
35) Dibromofluoromethane	5.391	113	63625	46.164	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.320%
50) Toluene-d8	8.647	98	231313	48.266	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.540%
62) 4-Bromofluorobenzene	11.079	95	90427	48.304	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110822\
Data File : VX032644.D
Acq On : 08 Nov 2022 11:21
Operator : JC/MD
Sample : VX1108WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1108WBL01

Quant Time: Nov 09 04:35:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

