

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110822\
 Data File : VX032648.D
 Acq On : 08 Nov 2022 12:57
 Operator : JC/MD
 Sample : N5484-19DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GES-1DL

Quant Time: Nov 09 04:37:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	116247	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	191742	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	171100	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83326	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	79712	49.350	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.700%
35) Dibromofluoromethane	5.385	113	62720	45.858	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.720%
50) Toluene-d8	8.647	98	232033	48.789	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.580%
62) 4-Bromofluorobenzene	11.079	95	90778	48.865	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.720%
Target Compounds						
					Qvalue	
31) Cyclohexane	5.477	56	4373	2.233	ug/l #	79
39) Methylcyclohexane	7.379	83	2919	1.295	ug/l #	90
40) Benzene	6.037	78	11925	2.213	ug/l	97
52) Toluene	8.720	92	42326	12.037	ug/l	97
67) Ethyl Benzene	10.195	91	359297	52.423	ug/l	99
68) m/p-Xylenes	10.305	106	46685	17.542	ug/l	100
69) o-Xylene	10.640	106	51516	19.909	ug/l	99
73) Isopropylbenzene	10.963	105	28714	4.349	ug/l	99
78) n-propylbenzene	11.305	91	72741	9.562	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	21962	3.892	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	94917	16.911	ug/l	100
85) sec-Butylbenzene	11.890	105	4603	0.679	ug/l	97
86) p-Isopropyltoluene	12.012	119	2303	0.400	ug/l	86
89) n-Butylbenzene	12.329	91	7124	1.424	ug/l #	69
95) Naphthalene	13.774	128	102643	15.691	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110822\
Data File : VX032648.D
Acq On : 08 Nov 2022 12:57
Operator : JC/MD
Sample : N5484-19DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GES-1DL

Quant Time: Nov 09 04:37:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

