

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110822\
 Data File : VX032661.D
 Acq On : 08 Nov 2022 17:59
 Operator : JC/MD
 Sample : N5415-08
 Misc : 5.04g/10ml/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 R-21

Quant Time: Nov 08 18:24:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	116456	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	189976	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	164541	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	80511	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	78762	48.675	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.340%
35) Dibromofluoromethane	5.385	113	57343	42.316	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	84.640%
50) Toluene-d8	8.647	98	229790	48.766	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.540%
62) 4-Bromofluorobenzene	11.085	95	78882	42.856	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110822\
Data File : VX032661.D
Acq On : 08 Nov 2022 17:59
Operator : JC/MD
Sample : N5415-08
Misc : 5.04g/10ml/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
R-21

Quant Time: Nov 08 18:24:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

