

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110822\
 Data File : VX032663.D
 Acq On : 08 Nov 2022 18:46
 Operator : JC/MD
 Sample : N5310-15
 Misc : 5.03g/10ml/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BIN 1-2-(2840)-VOC-11

Quant Time: Nov 09 04:41:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	109341	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	179928	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	157031	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	74724	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	75021	49.380	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.760%
35) Dibromofluoromethane	5.385	113	53840	41.950	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	83.900%
50) Toluene-d8	8.647	98	214207	47.998	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.000%
62) 4-Bromofluorobenzene	11.085	95	73510	42.168	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	84.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110822\
Data File : VX032663.D
Acq On : 08 Nov 2022 18:46
Operator : JC/MD
Sample : N5310-15
Misc : 5.03g/10ml/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BIN 1-2-(2840)-VOC-11

Quant Time: Nov 09 04:41:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

