

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX110822\
 Data File : VX032642.D
 Acq On : 08 Nov 2022 10:28
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 09 04:33:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	87	0.00
2 T	Dichlorodifluoromethane	50.000	47.986	4.0	79	0.00
3 P	Chloromethane	50.000	45.371	9.3	77	0.00
4 C	Vinyl Chloride	50.000	47.217	5.6#	79	0.00
5 T	Bromomethane	50.000	50.867	-1.7	86	0.00
6 T	Chloroethane	50.000	62.513	-25.0#	115	0.00
7 T	Trichlorofluoromethane	50.000	50.068	-0.1	87	0.00
8 T	Diethyl Ether	50.000	49.539	0.9	88	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	50.653	-1.3	91	0.00
10 T	Methyl Iodide	50.000	46.061	7.9	80	0.00
11 T	Tert butyl alcohol	250.000	200.808	19.7	71	0.02
12 CM	1,1-Dichloroethene	50.000	48.139	3.7#	85	0.00
13 T	Acrolein	250.000	244.852	2.1	89	0.00
14 T	Allyl chloride	50.000	50.497	-1.0	88	0.00
15 T	Acrylonitrile	250.000	228.521	8.6	83	0.00
16 T	Acetone	250.000	258.462	-3.4	98	0.00
17 T	Carbon Disulfide	50.000	44.488	11.0	76	0.00
18 T	Methyl Acetate	50.000	50.557	-1.1	89	0.00
19 T	Methyl tert-butyl Ether	50.000	51.376	-2.8	93	0.00
20 T	Methylene Chloride	50.000	47.770	4.5	90	0.00
21 T	trans-1,2-Dichloroethene	50.000	48.038	3.9	86	0.00
22 T	Diisopropyl ether	50.000	51.999	-4.0	94	0.00
23 T	Vinyl Acetate	250.000	251.561	-0.6	88	0.00
24 P	1,1-Dichloroethane	50.000	50.787	-1.6	91	0.00
25 T	2-Butanone	250.000	237.442	5.0	86	0.00
26 T	2,2-Dichloropropane	50.000	56.257	-12.5	98	0.00
27 T	cis-1,2-Dichloroethene	50.000	50.302	-0.6	88	0.00
28 T	Bromochloromethane	50.000	46.974	6.1	88	0.00
29 T	Tetrahydrofuran	250.000	227.383	9.0	82	0.00
30 C	Chloroform	50.000	52.315	-4.6#	94	0.00
31 T	Cyclohexane	50.000	49.340	1.3	85	0.00
32 T	1,1,1-Trichloroethane	50.000	53.212	-6.4	93	0.00
33 S	1,2-Dichloroethane-d4	50.000	50.183	-0.4	92	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	87	0.00
35 S	Dibromofluoromethane	50.000	51.389	-2.8	92	0.00
36 T	1,1-Dichloropropene	50.000	50.823	-1.6	89	0.00
37 T	Ethyl Acetate	50.000	45.775	8.5	85	0.00
38 T	Carbon Tetrachloride	50.000	53.341	-6.7	91	0.00
39 T	Methylcyclohexane	50.000	50.216	-0.4	84	0.00
40 TM	Benzene	50.000	51.312	-2.6	89	0.00
41 T	Methacrylonitrile	50.000	51.539	-3.1	90	0.00
42 TM	1,2-Dichloroethane	50.000	53.130	-6.3	92	0.00
43 T	Isopropyl Acetate	50.000	51.003	-2.0	87	0.00
44 TM	Trichloroethene	50.000	51.466	-2.9	91	0.00
45 C	1,2-Dichloropropane	50.000	52.468	-4.9#	92	0.00
46 T	Dibromomethane	50.000	51.322	-2.6	90	0.00
47 T	Bromodichloromethane	50.000	53.893	-7.8	93	0.00
48 T	Methyl methacrylate	50.000	49.569	0.9	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110822\
 Data File : VX032642.D
 Acq On : 08 Nov 2022 10:28
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 09 04:33:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	853.626	14.6	73	0.01
50 S	Toluene-d8	50.000	50.179	-0.4	88	0.00
51 T	4-Methyl-2-Pentanone	250.000	242.987	2.8	84	0.00
52 CM	Toluene	50.000	52.225	-4.5#	90	0.00
53 T	t-1,3-Dichloropropene	50.000	56.455	-12.9	94	0.00
54 T	cis-1,3-Dichloropropene	50.000	55.383	-10.8	94	0.00
55 T	1,1,2-Trichloroethane	50.000	51.892	-3.8	93	0.00
56 T	Ethyl methacrylate	50.000	53.220	-6.4	89	0.00
57 T	1,3-Dichloropropane	50.000	51.922	-3.8	91	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	262.883	-5.2	88	0.00
59 T	2-Hexanone	250.000	249.194	0.3	85	0.00
60 T	Dibromochloromethane	50.000	53.924	-7.8	93	0.00
61 T	1,2-Dibromoethane	50.000	52.471	-4.9	91	0.00
62 S	4-Bromofluorobenzene	50.000	53.206	-6.4	92	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	86	0.00
64 T	Tetrachloroethene	50.000	51.831	-3.7	91	0.00
65 PM	Chlorobenzene	50.000	51.945	-3.9	92	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	54.454	-8.9	95	0.00
67 C	Ethyl Benzene	50.000	53.318	-6.6#	92	0.00
68 T	m/p-Xylenes	100.000	106.123	-6.1	91	0.00
69 T	o-Xylene	50.000	53.701	-7.4	92	0.00
70 T	Styrene	50.000	54.601	-9.2	93	0.00
71 P	Bromoform	50.000	54.221	-8.4	91	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	86	0.00
73 T	Isopropylbenzene	50.000	51.348	-2.7	93	0.00
74 T	N-ethyl acetate	50.000	51.690	-3.4	89	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	48.025	4.0	88	0.00
76 T	1,2,3-Trichloropropane	50.000	44.126	11.7	78	0.00
77 T	Bromobenzene	50.000	50.215	-0.4	92	0.00
78 T	n-propylbenzene	50.000	52.316	-4.6	93	0.00
79 T	2-Chlorotoluene	50.000	51.391	-2.8	92	0.00
80 T	1,3,5-Trimethylbenzene	50.000	52.693	-5.4	92	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	50.927	-1.9	87	0.00
82 T	4-Chlorotoluene	50.000	52.296	-4.6	93	0.00
83 T	tert-Butylbenzene	50.000	51.403	-2.8	91	0.00
84 T	1,2,4-Trimethylbenzene	50.000	51.994	-4.0	91	0.00
85 T	sec-Butylbenzene	50.000	52.469	-4.9	91	0.00
86 T	p-Isopropyltoluene	50.000	53.209	-6.4	91	0.00
87 T	1,3-Dichlorobenzene	50.000	51.413	-2.8	92	0.00
88 T	1,4-Dichlorobenzene	50.000	50.860	-1.7	93	0.00
89 T	n-Butylbenzene	50.000	53.313	-6.6	91	0.00
90 T	Hexachloroethane	50.000	53.582	-7.2	93	0.00
91 T	1,2-Dichlorobenzene	50.000	52.039	-4.1	93	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	46.181	7.6	80	0.00
93 T	1,2,4-Trichlorobenzene	50.000	50.272	-0.5	92	0.00
94 T	Hexachlorobutadiene	50.000	50.285	-0.6	93	0.00
95 T	Naphthalene	50.000	48.100	3.8	85	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110822\
Data File : VX032642.D
Acq On : 08 Nov 2022 10:28
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Nov 09 04:33:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	49.093	1.8	90	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6