

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110822\
 Data File : VX032664.D
 Acq On : 08 Nov 2022 19:09
 Operator : JC/MD
 Sample : N5415-01
 Misc : 5.00g/10ml/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 R-14

Quant Time: Nov 09 04:41:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	111715	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	187388	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	162261	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	75796	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	78014	50.259	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.520%
35) Dibromofluoromethane	5.391	113	56059	41.940	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	83.880%
50) Toluene-d8	8.647	98	223117	48.004	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.000%
62) 4-Bromofluorobenzene	11.085	95	76330	42.042	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	84.080%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110822\
Data File : VX032664.D
Acq On : 08 Nov 2022 19:09
Operator : JC/MD
Sample : N5415-01
Misc : 5.00g/10ml/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
R-14

Quant Time: Nov 09 04:41:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 711
Delta R.T. -0.006 min
Lab File: VX032664.D
Acq: 08 Nov 2022 19:09




Instrument :
MSVOA_X
ClientSampleId :
R-14


Tgt Ion:168 Resp: 111715
Ion Ratio Lower Upper
168 100
99 60.9 46.1 69.1



#33
1,2-Dichloroethane-d4
Concen: 50.259 ug/l
RT: 5.958 min Scan# 799
Delta R.T. -0.000 min
Lab File: VX032664.D
Acq: 08 Nov 2022 19:09

Tgt Ion: 65 Resp: 78014
Ion Ratio Lower Upper
65 100
67 50.9 0.0 101.8









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 91
Delta R.T. -0.000 min
Lab File: VX032664.D
Acq: 08 Nov 2022 19:09




Instrument :
MSVOA_X
ClientSampleId :
R-14


Tgt Ion:114 Resp: 187388
Ion Ratio Lower Upper
114 100
63 19.0 0.0 41.6
88 15.5 0.0 33.4



#35
Dibromofluoromethane
Concen: 41.940 ug/l
RT: 5.391 min Scan# 706
Delta R.T. 0.006 min
Lab File: VX032664.D
Acq: 08 Nov 2022 19:09

Tgt Ion:113 Resp: 56059
Ion Ratio Lower Upper
113 100
111 102.1 82.4 123.6
192 19.5 15.4 23.2









#50
Toluene-d8
Concen: 48.004 ug/l
RT: 8.647 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX032664.D
Acq: 08 Nov 2022 19:09




Instrument :
MSVOA_X
ClientSampleId :
R-14


Tgt Ion: 98 Resp: 223117
Ion Ratio Lower Upper
98 100
100 64.5 50.9 76.3



#62
4-Bromofluorobenzene
Concen: 42.042 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.006 min
Lab File: VX032664.D
Acq: 08 Nov 2022 19:09

Tgt Ion: 95 Resp: 76330
Ion Ratio Lower Upper
95 100
174 82.0 0.0 164.0
176 80.2 0.0 154.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX032664.D
Acq: 08 Nov 2022 19:09

Instrument :
MSVOA_X
ClientSampleId :
R-14

Tgt Ion:117 Resp: 162261
Ion Ratio Lower Upper
117 100
82 55.1 44.3 66.5
119 30.8 25.7 38.5



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. -0.000 min
Lab File: VX032664.D
Acq: 08 Nov 2022 19:09

Tgt Ion:152 Resp: 75796
Ion Ratio Lower Upper
152 100
115 61.7 44.0 132.0
150 157.4 0.0 351.8

