

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110823\
 Data File : VX038643.D
 Acq On : 08 Nov 2023 09:41
 Operator : JC/MD
 Sample : VX1108MBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1108MBL01

Quant Time: Nov 09 01:17:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 08 03:19:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	290374	50.000	ug/l	#-0.01
34) 1,4-Difluorobenzene	6.757	114	475379	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	484206	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	265457	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	164385	45.472	ug/l	-0.01
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.940%
35) Dibromofluoromethane	5.379	113	153905	44.951	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.900%
50) Toluene-d8	8.647	98	561160	45.802	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.600%#
62) 4-Bromofluorobenzene	11.079	95	236966	49.886	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.780%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110823\
Data File : VX038643.D
Acq On : 08 Nov 2023 09:41
Operator : JC/MD
Sample : VX1108MBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1108MBL01

Quant Time: Nov 09 01:17:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 08 03:19:07 2023
Response via : Initial Calibration

