

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110823\
 Data File : VX038648.D
 Acq On : 08 Nov 2023 12:12
 Operator : JC/MD
 Sample : 05253-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-12

Quant Time: Nov 09 01:19:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 08 03:19:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	274602	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	449188	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	467444	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	262029	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	156267	45.709	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.420%
35) Dibromofluoromethane	5.385	113	144827	44.766	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.540%
50) Toluene-d8	8.647	98	532490	45.996	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.000%
62) 4-Bromofluorobenzene	11.079	95	226896	50.551	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.100%
Target Compounds						
					Qvalue	
16) Acetone	2.373	43	3107	2.158	ug/l	# 87
17) Carbon Disulfide	2.514	76	1970	0.289	ug/l	98
20) Methylene Chloride	2.794	84	1235	0.426	ug/l	# 74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110823\
Data File : VX038648.D
Acq On : 08 Nov 2023 12:12
Operator : JC/MD
Sample : 05253-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-12

Quant Time: Nov 09 01:19:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 08 03:19:07 2023
Response via : Initial Calibration

