

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110823\
 Data File : VX038657.D
 Acq On : 08 Nov 2023 15:40
 Operator : JC/MD
 Sample : 05222-09RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-MW-EVAL-02D-20231101RE

Quant Time: Nov 09 01:20:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 08 03:19:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	279908	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	459695	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	464455	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	265705	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	154454	44.323	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.640%
35) Dibromofluoromethane	5.385	113	144292	43.581	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.160%
50) Toluene-d8	8.647	98	519967	43.888	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	87.780%#
62) 4-Bromofluorobenzene	11.079	95	219107	47.700	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.400%
Target Compounds						
					Qvalue	
16) Acetone	2.374	43	3194	2.176	ug/l	# 73
20) Methylene Chloride	2.788	84	869	0.294	ug/l	# 80
25) 2-Butanone	4.581	43	4726	2.223	ug/l	100
44) Trichloroethene	7.129	130	2114	0.574	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110823\
Data File : VX038657.D
Acq On : 08 Nov 2023 15:40
Operator : JC/MD
Sample : 05222-09RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-MW-EVAL-02D-20231101RE

Quant Time: Nov 09 01:20:44 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 08 03:19:07 2023
Response via : Initial Calibration

