

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110823\
 Data File : VX038658.D
 Acq On : 08 Nov 2023 16:04
 Operator : JC/MD
 Sample : 05223-13RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-MW-20-20231030RE

Quant Time: Nov 09 01:20:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 08 03:19:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	285082	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	477315	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	488151	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	274536	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	161130	45.399	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.800%
35) Dibromofluoromethane	5.385	113	149467	43.478	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	86.960%
50) Toluene-d8	8.647	98	543707	44.198	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	88.400%#
62) 4-Bromofluorobenzene	11.079	95	228522	47.913	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.820%
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.788	84	1032	0.342	ug/l	# 74
44) Trichloroethene	7.129	130	178034	46.521	ug/l	90
64) Tetrachloroethene	9.275	164	5853	1.602	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110823\
Data File : VX038658.D
Acq On : 08 Nov 2023 16:04
Operator : JC/MD
Sample : 05223-13RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-MW-20-20231030RE

Quant Time: Nov 09 01:20:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 08 03:19:07 2023
Response via : Initial Calibration

