

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110823\
 Data File : VX038660.D
 Acq On : 08 Nov 2023 16:50
 Operator : JC/MD
 Sample : 05223-20RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-MW-18-20231030RE

Quant Time: Nov 09 01:21:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 08 03:19:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	267806	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	447408	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	458211	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	257417	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	150821	45.236	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.480%
35) Dibromofluoromethane	5.385	113	141464	43.901	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.800%
50) Toluene-d8	8.647	98	505113	43.805	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	87.620%#
62) 4-Bromofluorobenzene	11.079	95	215679	48.243	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.480%
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.788	84	942	0.333	ug/l	# 77
32) 1,1,1-Trichloroethane	5.391	97	7253	1.574	ug/l	96
44) Trichloroethene	7.123	130	209776	58.479	ug/l	94
64) Tetrachloroethene	9.281	164	2333	0.680	ug/l	# 75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110823\
Data File : VX038660.D
Acq On : 08 Nov 2023 16:50
Operator : JC/MD
Sample : 05223-20RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-MW-18-20231030RE

Quant Time: Nov 09 01:21:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 08 03:19:07 2023
Response via : Initial Calibration

