

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110823\
 Data File : VX038661.D
 Acq On : 08 Nov 2023 17:13
 Operator : JC/MD
 Sample : 05223-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-B-3-20231031

Quant Time: Nov 09 01:21:34 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 08 03:19:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	264205	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	432553	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	447694	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	253291	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	153872	46.780	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.560%
35) Dibromofluoromethane	5.385	113	142956	45.887	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.780%
50) Toluene-d8	8.647	98	514797	46.178	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.360%
62) 4-Bromofluorobenzene	11.079	95	217737	50.376	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.760%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	3376	2.437	ug/l	91
20) Methylene Chloride	2.794	84	732	0.262	ug/l	# 79
44) Trichloroethene	7.135	130	960	0.277	ug/l	57
64) Tetrachloroethene	9.275	164	19513	5.823	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110823\
Data File : VX038661.D
Acq On : 08 Nov 2023 17:13
Operator : JC/MD
Sample : 05223-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-B-3-20231031

Quant Time: Nov 09 01:21:34 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 08 03:19:07 2023
Response via : Initial Calibration

