

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110823\
 Data File : VX038664.D
 Acq On : 08 Nov 2023 18:22
 Operator : JC/MD
 Sample : 05223-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-MW-36-20231030

Quant Time: Nov 09 01:22:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 08 03:19:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	276271	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	453257	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	478476	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	288261	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	160514	46.668	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.340%
35) Dibromofluoromethane	5.385	113	155574	47.656	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.320%
50) Toluene-d8	8.647	98	526175	45.043	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	90.080%#
62) 4-Bromofluorobenzene	11.079	95	242362	53.512	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.020%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.953	59	1575	2.240	ug/l	# 87
16) Acetone	2.380	43	2788	1.925	ug/l	# 80
20) Methylene Chloride	2.794	84	973	0.333	ug/l	# 72
44) Trichloroethene	7.129	130	2499	0.688	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110823\
Data File : VX038664.D
Acq On : 08 Nov 2023 18:22
Operator : JC/MD
Sample : 05223-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-MW-36-20231030

Quant Time: Nov 09 01:22:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 08 03:19:07 2023
Response via : Initial Calibration

