

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110823\
 Data File : VX038666.D
 Acq On : 08 Nov 2023 19:08
 Operator : JC/MD
 Sample : 05223-19
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-FD-1-20231030

Quant Time: Nov 09 01:22:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 08 03:19:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	278629	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	459020	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	483462	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	285314	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	162580	46.869	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.740%
35) Dibromofluoromethane	5.385	113	156506	47.340	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.680%
50) Toluene-d8	8.647	98	535534	45.268	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	90.540%#
62) 4-Bromofluorobenzene	11.079	95	241718	52.700	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.400%
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.788	84	802	0.272	ug/l	# 74
32) 1,1,1-Trichloroethane	5.385	97	15373	3.206	ug/l	94
44) Trichloroethene	7.129	130	23864	6.484	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110823\
Data File : VX038666.D
Acq On : 08 Nov 2023 19:08
Operator : JC/MD
Sample : 05223-19
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-FD-1-20231030

Quant Time: Nov 09 01:22:41 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 08 03:19:07 2023
Response via : Initial Calibration

