

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110823\
 Data File : VX038657.D
 Acq On : 08 Nov 2023 15:40
 Operator : JC/MD
 Sample : 05222-09RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-MW-EVAL-02D-20231101RE

Quant Time: Nov 09 01:20:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 08 03:19:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	5.556	168	279908	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	459695	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	464455	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	265705	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	154454	44.323	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.640%
35) Dibromofluoromethane	5.385	113	144292	43.581	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.160%
50) Toluene-d8	8.647	98	519967	43.888	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	87.780%#
62) 4-Bromofluorobenzene	11.079	95	219107	47.700	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.400%
Target Compounds						
					Qvalue	
16) Acetone	2.374	43	3194	2.176	ug/l	# 73
20) Methylene Chloride	2.788	84	869	0.294	ug/l	# 80
25) 2-Butanone	4.581	43	4726	2.223	ug/l	100
44) Trichloroethene	7.129	130	2114	0.574	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110823\
Data File : VX038657.D
Acq On : 08 Nov 2023 15:40
Operator : JC/MD
Sample : 05222-09RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-MW-EVAL-02D-20231101RE

Quant Time: Nov 09 01:20:44 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 08 03:19:07 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX038657.D
Acq: 08 Nov 2023 15:40




Instrument :
MSVOA_X
ClientSampleId :
FSND-MW-EVAL-02D-20231101RE


Tgt Ion:168 Resp: 279908
Ion Ratio Lower Upper
168 100
99 52.4 56.6 85.0#



#16
Acetone
Concen: 2.176 ug/l
RT: 2.374 min Scan# 211
Delta R.T. -0.006 min
Lab File: VX038657.D
Acq: 08 Nov 2023 15:40

Tgt Ion: 43 Resp: 3194
Ion Ratio Lower Upper
43 100
58 41.3 21.9 32.9#









#20
Methylene Chloride
Concen: 0.294 ug/l
RT: 2.788 min Scan# 21
Delta R.T. 0.000 min
Lab File: VX038657.D
Acq: 08 Nov 2023 15:40




Instrument :
MSVOA_X
ClientSampleId :
FSND-MW-EVAL-02D-20231101RE


Tgt Ion:	84	Resp:	869
Ion	Ratio	Lower	Upper
84	100		
49	99.6	103.6	155.4#
51	47.5	33.8	50.6
86	49.0	49.4	74.0#



#25
2-Butanone
Concen: 2.223 ug/l
RT: 4.581 min Scan# 573
Delta R.T. 0.025 min
Lab File: VX038657.D
Acq: 08 Nov 2023 15:40

Tgt Ion:	43	Resp:	4726
Ion	Ratio	Lower	Upper
43	100		
72	24.4	19.4	29.2









#33
1,2-Dichloroethane-d4
Concen: 44.323 ug/l
RT: 5.952 min Scan# 799
Delta R.T. -0.006 min
Lab File: VX038657.D
Acq: 08 Nov 2023 15:40




Instrument : MSVOA_X
ClientSampleId : FSND-MW-EVAL-02D-20231101RE


Tgt Ion: 65 Resp: 154454
Ion Ratio Lower Upper
65 100
67 51.6 0.0 100.2



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 931
Delta R.T. 0.000 min
Lab File: VX038657.D
Acq: 08 Nov 2023 15:40

Tgt Ion:114 Resp: 459695
Ion Ratio Lower Upper
114 100
63 17.2 0.0 45.0
88 15.4 0.0 36.6









#35
Dibromofluoromethane
Concen: 43.581 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.000 min
Lab File: VX038657.D
Acq: 08 Nov 2023 15:40




Instrument :
MSVOA_X
ClientSampleId :
FSND-MW-EVAL-02D-20231101RE


Tgt Ion:113 Resp: 144292
Ion Ratio Lower Upper
113 100
111 101.5 82.9 124.3
192 20.9 13.3 19.9#



#44
Trichloroethene
Concen: 0.574 ug/l
RT: 7.129 min Scan# 991
Delta R.T. 0.000 min
Lab File: VX038657.D
Acq: 08 Nov 2023 15:40

Tgt Ion:130 Resp: 2114
Ion Ratio Lower Upper
130 100
95 99.6 0.0 204.4









#50
Toluene-d8
Concen: 43.888 ug/l
RT: 8.647 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX038657.D
Acq: 08 Nov 2023 15:40




Instrument :
MSVOA_X
ClientSampleId :
FSND-MW-EVAL-02D-20231101RE


Tgt Ion: 98 Resp: 519967
Ion Ratio Lower Upper
98 100
100 64.6 53.1 79.7



#62
4-Bromofluorobenzene
Concen: 47.700 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX038657.D
Acq: 08 Nov 2023 15:40

Tgt Ion: 95 Resp: 219107
Ion Ratio Lower Upper
95 100
174 87.3 0.0 135.6
176 84.8 0.0 131.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX038657.D
Acq: 08 Nov 2023 15:40

Instrument :
MSVOA_X
ClientSampleId :
FSND-MW-EVAL-02D-20231101RE

Tgt Ion:117 Resp: 464455
Ion Ratio Lower Upper
117 100
82 49.7 46.2 69.4
119 32.7 25.5 38.3



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX038657.D
Acq: 08 Nov 2023 15:40

Tgt Ion:152 Resp: 265705
Ion Ratio Lower Upper
152 100
115 57.6 43.3 129.9
150 156.3 0.0 345.0

