



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110823\
Data File : VX038641.D
Acq On : 08 Nov 2023 08:10
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
Title : SW846 8260
Last Update : Wed Nov 08 03:19:07 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.9	16673	PASS
75	95	30	60	50.0	52389	PASS
95	95	100	100	100.0	104805	PASS
96	95	5	9	7.0	7307	PASS
173	174	0.00	2	0.8	677	PASS
174	95	50	100	85.7	89861	PASS
175	174	5	9	7.7	6934	PASS
176	174	95	101	96.9	87080	PASS
177	176	5	9	6.8	5878	PASS