

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110824\
 Data File : VX043773.D
 Acq On : 08 Nov 2024 18:40
 Operator : JC/MD
 Sample : P4742-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 E1H26

Quant Time: Nov 09 00:01:57 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML110424WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Nov 08 23:56:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.757	114	357224	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	322735	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	141279	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	88135	46.312	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery = 92.620%			
7) Chloroethane-d5	1.648	69	45908	56.121	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery = 112.240%			
11) 1,1-Dichloroethene-d2	2.294	65	38976	44.366	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 88.740%			
21) 2-Butanone-d5	4.459	46	113084	90.755	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery = 90.760%			
24) Chloroform-d	5.050	84	178923	43.613	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 87.220%			
26) 1,2-Dichloroethane-d4	5.946	65	115195	44.200	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 88.400%			
32) Benzene-d6	5.964	84	378092	45.313	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 90.620%			
36) 1,2-Dichloropropane-d6	7.306	67	115485	45.492	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery = 90.980%			
41) Toluene-d8	8.647	98	339243	44.756	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 89.520%			
43) trans-1,3-Dichloroprop...	8.952	79	46502	38.304	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 76.600%			
47) 2-Hexanone-d5	9.384	63	91175	90.754	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery = 90.750%			
56) 1,1,2,2-Tetrachloroeth...	11.189	84	156549	48.064	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery = 96.120%			
66) 1,2-Dichlorobenzene-d4	12.317	152	128043	52.309	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 104.620%			
Target Compounds						Qvalue
34) Trichloroethene	7.117	95	489023	197.736	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110824\
Data File : VX043773.D
Acq On : 08 Nov 2024 18:40
Operator : JC/MD
Sample : P4742-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
E1H26

Quant Time: Nov 09 00:01:57 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM110424WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Nov 08 23:56:14 2024
Response via : Initial Calibration

