

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110918\
 Data File : VX005862.D
 Acq On : 09 Nov 2018 15:32
 Operator : JC/MD
 Sample : PB114659ZHE#05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB114659ZHE#05

Quant Time: Nov 11 23:51:00 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	109353	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	6.86	114	153024	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	142696	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	67136	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.06	65	71252	56.68	ug/l	0.00
Spiked Amount	50.000		Recovery	=	113.36%	
35) Dibromofluoromethane	5.50	113	58113	55.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	111.68%	
50) Toluene-d8	8.71	98	191467	55.33	ug/l	0.00
Spiked Amount	50.000		Recovery	=	110.66%	
62) 4-Bromofluorobenzene	11.14	95	64237	51.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.26%	

Target Compounds


						Qvalue
3) Chloromethane	1.32	50	314	0.243	ug/l #	87
6) Chloroethane	1.73	64	180	0.232	ug/l #	42
11) Tert butyl alcohol	3.03	59	259	0.719	ug/l #	72
16) Acetone	2.44	43	6461	9.545	ug/l	92
18) Methyl Acetate	2.78	43	10035	4.173	ug/l	99
20) Methylene Chloride	2.86	84	1407	1.265	ug/l	88
25) 2-Butanone	4.71	43	870	0.818	ug/l #	90
29) Tetrahydrofuran	5.15	42	475	0.695	ug/l #	36
31) Cyclohexane	5.67	56	2057	1.175	ug/l #	1
36) 1,1-Dichloropropene	5.67	75	7488	5.076	ug/l #	49
38) Carbon Tetrachloride	5.66	117	9939	6.730	ug/l #	16
41) Methacrylonitrile	5.17	41	223	0.213	ug/l #	47
43) Isopropyl Acetate	6.46	43	26249	9.338	ug/l	98
48) Methyl methacrylate	7.68	41	8441	6.688	ug/l #	16
49) 1,4-Dioxane	7.88	88	115	4.228	ug/l #	1
51) 4-Methyl-2-Pentanone	8.62	43	62175	35.096	ug/l #	51
54) cis-1,3-Dichloropropene	8.62	75	27600	18.537	ug/l #	62
56) Ethyl methacrylate	9.30	69	689	0.430	ug/l #	1
57) 1,3-Dichloropropane	9.54	76	6947	3.859	ug/l #	43
59) 2-Hexanone	9.54	43	102863	65.872	ug/l #	52
74) N-amyl acetate	10.87	43	360	1.412	ug/l #	44
76) 1,2,3-Trichloropropane	11.14	75	32279	23.612	ug/l #	36
80) 1,3,5-Trimethylbenzene	11.51	105	5854	1.774	ug/l	99
81) trans-1,4-Dichloro-2-buten	11.14	75	32279	75.257	ug/l #	9
82) 4-Chlorotoluene	11.51	91	737	0.226	ug/l #	44
83) tert-Butylbenzene	11.51	119	793	0.238	ug/l #	52
84) 1,2,4-Trimethylbenzene	11.81	105	170	Below Cal		72
95) Naphthalene	13.84	128	136	1.205	ug/l #	69


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110918\
Data File : VX005862.D
Acq On : 09 Nov 2018 15:32
Operator : JC/MD
Sample : PB114659ZHE#05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB114659ZHE#05

Quant Time: Nov 11 23:51:00 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.67 min Scan# 741
Delta R.T. 0.01 min
Lab File: VX005862.D
Acq: 09 Nov 2018 15:32




Instrument :
MSVOA_X
ClientSampleId :
PB114659ZHE#05


Tot Ion:168 Resp: 109353
Ion Ratio Lower Upper
168 100
99 45.0 40.7 61.1



#3
Chloromethane
Concen: 0.243 ug/l
RT: 1.32 min Scan# 28
Delta R.T. -0.00 min
Lab File: VX005862.D
Acq: 09 Nov 2018 15:32

Tgt Ion: 50 Resp: 314
Ion Ratio Lower Upper
50 100
52 40.2 26.2 39.2#









#6
Chloroethane
Concen: 0.232 ug/l
RT: 1.73 min Scan# 94
Delta R.T. 0.01 min
Lab File: VX005862.D
Acq: 09 Nov 2018 15:32




Instrument :
MSVOA_X
ClientSampleId :
PB114659ZHE#05


Tot Ion: 64 Resp: 180
Ion Ratio Lower Upper
64 100
66 0.0 25.8 38.6#



#11
Tert butyl alcohol
Concen: 0.719 ug/l
RT: 3.03 min Scan# 307
Delta R.T. -0.05 min
Lab File: VX005862.D
Acq: 09 Nov 2018 15:32

Tgt Ion: 59 Resp: 259
Ion Ratio Lower Upper
59 100
57 0.0 8.5 12.7#





#16

Acetone

Concen: 9.545 ug/l

RT: 2.44 min Scan# 211

Delta R.T. -0.01 min

Lab File: VX005862.D

Acq: 09 Nov 2018 15:32

Instrument :

MSVOA_X

ClientSampleId :


PB114659ZHE#05

Tot Ion: 43 Resp: 6461

Ion Ratio Lower Upper




43 100

58 25.1 23.8 35.6



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0



#18

Methyl Acetate

Concen: 4.173 ug/l

RT: 2.78 min Scan# 266

Delta R.T. -0.01 min

Lab File: VX005862.D


Acq: 09 Nov 2018 15:32

Tgt Ion: 43 Resp: 10035

Ion Ratio Lower Upper



43 100


74 20.7 16.8 25.2



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 74.00 (73.70 to 74.70): VX0









#20
Methvlene Chloride
Concen: 1.265 ug/l
RT: 2.86 min Scan# 280
Delta R.T. 0.01 min
Lab File: VX005862.D
Acq: 09 Nov 2018 15:32




Instrument :
MSVOA_X
ClientSampleId :
PB114659ZHE#05


Tot Ion:	84	Resp:	1407
Ion	Ratio	Lower	Upper
84	100		
49	121.4	109.5	164.3
51	48.6	33.3	49.9
86	56.8	52.7	79.1



#25
2-Butanone
Concen: 0.818 ug/l
RT: 4.71 min Scan# 584
Delta R.T. 0.01 min
Lab File: VX005862.D
Acq: 09 Nov 2018 15:32

Tgt Ion:	43	Resp:	870
Ion	Ratio	Lower	Upper
43	100		
72	19.2	19.2	28.8#









#29
Tetrahydrofuran
Concen: 0.695 ug/l
RT: 5.15 min Scan# 656
Delta R.T. -0.00 min
Lab File: VX005862.D
Acq: 09 Nov 2018 15:32




Instrument :
MSVOA_X
ClientSampleId :
PB114659ZHE#05


Tot Ion	Ratio	Lower	Upper
42	100		
72	0.0	32.0	48.0#
71	0.0	29.8	44.8#



#31
Cyclohexane
Concen: 1.175 ug/l
RT: 5.67 min Scan# 740
Delta R.T. 0.09 min
Lab File: VX005862.D
Acq: 09 Nov 2018 15:32

Tgt Ion	Ratio	Lower	Upper
56	100		
69	172.0	22.6	34.0#
84	154.8	62.7	94.1#









#33
 1,2-Dichloroethane-d4
 Concen: 56.677 ug/l
 RT: 6.06 min Scan# 805
 Delta R.T. -0.01 min
 Lab File: VX005862.D
 Acq: 09 Nov 2018 15:32




Instrument :
 MSVOA_X
 ClientSampleId :
 PB114659ZHE#05


Tot Ion: 65 Resp: 71252
 Ion Ratio Lower Upper
 65 100
 67 52.6 0.0 105.0



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.86 min Scan# 936
 Delta R.T. -0.00 min
 Lab File: VX005862.D
 Acq: 09 Nov 2018 15:32

Tgt Ion: 114 Resp: 153024
 Ion Ratio Lower Upper
 114 100
 63 20.2 0.0 42.8
 88 14.9 0.0 31.2









#35
 Dibromofluoromethane
 Concen: 55.835 ug/l
 RT: 5.50 min Scan# 713
 Delta R.T. -0.01 min
 Lab File: VX005862.D
 Acq: 09 Nov 2018 15:32




Instrument :
 MSVOA_X
 ClientSampleId :
 PB114659ZHE#05


Tot Ion:113 Resp: 58113
 Ion Ratio Lower Upper
 113 100
 111 98.7 82.3 123.5
 192 21.8 17.9 26.9



#36
 1,1-Dichloropropene
 Concen: 5.076 ug/l
 RT: 5.67 min Scan# 740
 Delta R.T. -0.13 min
 Lab File: VX005862.D
 Acq: 09 Nov 2018 15:32

Tgt Ion: 75 Resp: 7488
 Ion Ratio Lower Upper
 75 100
 110 9.3 17.4 52.2#
 77 0.0 25.8 38.6#





#38

Carbon Tetrachloride

Concen: 6.730 ug/l

RT: 5.66 min Scan# 739

Delta R.T. -0.12 min

Lab File: VX005862.D

Acq: 09 Nov 2018 15:32

Instrument :

MSVOA_X

ClientSampleId :

PB114659ZHE#05


Tot Ion:117 Resp: 9939

Ion Ratio Lower Upper

117 100

119 6.3 78.1 117.1#




121 0.0 24.6 36.8#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 118.80 (118.50 to 119.50): V

Ion 120.80 (120.50 to 121.50): V



#41

Methacrylonitrile

Concen: 0.213 ug/l

RT: 5.17 min Scan# 658

Delta R.T. 0.12 min

Lab File: VX005862.D

Acq: 09 Nov 2018 15:32

Tgt Ion: 41 Resp: 223


Ion Ratio Lower Upper

41 100

39 61.9 42.2 63.2

67 0.0 53.4 80.0#

52 0.0 23.4 35.2#






Abundance Ion 41.00 (40.70 to 41.70): VX0

Ion 39.00 (38.70 to 39.70): VX0

Ion 67.00 (66.70 to 67.70): VX0

Ion 52.00 (51.70 to 52.70): VX0





#43

Isopropyl Acetate

Concen: 9.338 ug/l

RT: 6.46 min Scan# 871

Delta R.T. -0.00 min

Lab File: VX005862.D

Acq: 09 Nov 2018 15:32

Instrument :


MSVOA_X

ClientSampleId :

PB114659ZHE#05

Tot Ion: 43 Resp: 26249

Ion	Ratio	Lower	Upper
43	100		
61	19.5	14.6	22.0
87	12.5	9.6	14.4






Abundance

Ion 43.00 (42.70 to 43.70): VX0

Ion 61.00 (60.70 to 61.70): VX0

Ion 87.00 (86.70 to 87.70): VX0



#48

Methyl methacrylate

Concen: 6.688 ug/l

RT: 7.68 min Scan# 1071


Delta R.T. -0.09 min

Lab File: VX005862.D

Acq: 09 Nov 2018 15:32

Tgt Ion: 41 Resp: 8441

Ion	Ratio	Lower	Upper
41	100		
69	0.0	63.2	94.8#
39	0.0	42.2	63.2#






Abundance

Ion 41.00 (40.70 to 41.70): VX0

Ion 69.00 (68.70 to 69.70): VX0

Ion 39.00 (38.70 to 39.70): VX0





#49

1,4-Dioxane

Concen: 4.228 ug/l

RT: 7.88 min Scan# 1104

Delta R.T. 0.13 min

Lab File: VX005862.D

Acq: 09 Nov 2018 15:32

Instrument :

MSVOA_X

ClientSampleId :

PB114659ZHE#05


Tot Ion: 88 Resp: 115

Ion Ratio Lower Upper

88 100

43 798430.4 27.4 41.0#




58 3257.4 59.4 89.2#



Abundance Ion 88.00 (87.70 to 88.70): VX0

Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0



#50

Toluene-d8

Concen: 55.334 ug/l

RT: 8.71 min Scan# 1240

Delta R.T. -0.00 min

Lab File: VX005862.D


Acq: 09 Nov 2018 15:32

Tgt Ion: 98 Resp: 191467

Ion Ratio Lower Upper



98 100


100 63.2 51.3 76.9



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): VX0





#51

4-Methyl-2-Pentanone

Concen: 35.096 ug/l

RT: 8.62 min Scan# 1224

Delta R.T. -0.04 min

Lab File: VX005862.D

Acq: 09 Nov 2018 15:32

Instrument :

MSVOA_X

ClientSampleId :


PB114659ZHE#05

Tot Ion: 43 Resp: 62175

Ion Ratio Lower Upper




43 100

58 7.9 29.7 44.5#



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0



#54

cis-1,3-Dichloropropene

Concen: 18.537 ug/l

RT: 8.62 min Scan# 1225

Delta R.T. 0.19 min

Lab File: VX005862.D

Acq: 09 Nov 2018 15:32


Tgt Ion: 75 Resp: 27600

Ion Ratio Lower Upper

75 100

77 0.6 25.2 37.8#



39 30.3 38.2 57.2#




Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 76.90 (76.60 to 77.60): VX0

Ion 39.00 (38.70 to 39.70): VX0





#56

Ethyl methacrylate

Concen: 0.430 ug/l

RT: 9.30 min Scan# 1337

Delta R.T. 0.12 min

Lab File: VX005862.D

Acq: 09 Nov 2018 15:32

Instrument :

MSVOA_X

ClientSampleId :

PB114659ZHE#05


Tot Ion: 69 Resp: 689

Ion Ratio Lower Upper

69 100

41 2774.7 57.0 85.6#




39 1123.1 28.3 42.5#



Abundance Ion 69.00 (68.70 to 69.70): VX0

Ion 41.00 (40.70 to 41.70): VX0

Ion 39.00 (38.70 to 39.70): VX0



#57

1,3-Dichloropropane

Concen: 3.859 ug/l

RT: 9.54 min Scan# 1376

Delta R.T. 0.17 min

Lab File: VX005862.D


Acq: 09 Nov 2018 15:32

Tgt Ion: 76 Resp: 6947

Ion Ratio Lower Upper



76 100


78 0.0 25.3 37.9#



Abundance Ion 75.90 (75.60 to 76.60): VX0

Ion 77.90 (77.60 to 78.60): VX0









#59
2-Hexanone
Concen: 65.872 ug/l
RT: 9.54 min Scan# 1376
Delta R.T. 0.04 min
Lab File: VX005862.D
Acq: 09 Nov 2018 15:32




Instrument :
MSVOA_X
ClientSampleId :
PB114659ZHE#05


Tot Ion: 43 Resp: 102863
Ion Ratio Lower Upper
43 100
58 18.2 25.9 77.8#



#62
4-Bromofluorobenzene
Concen: 51.635 ug/l
RT: 11.14 min Scan# 1638
Delta R.T. -0.00 min
Lab File: VX005862.D
Acq: 09 Nov 2018 15:32

Tgt Ion: 95 Resp: 64237
Ion Ratio Lower Upper
95 100
174 91.9 0.0 184.4
176 88.7 0.0 177.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.12 min Scan# 1470
Delta R.T. -0.00 min
Lab File: VX005862.D
Acq: 09 Nov 2018 15:32

Instrument :
MSVOA_X
ClientSampleId :
PB114659ZHE#05

Tot Ion:117 Resp: 142696
Ion Ratio Lower Upper
117 100
82 54.8 44.1 66.1
119 32.4 26.2 39.2




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V






#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.08 min Scan# 1792
Delta R.T. -0.00 min
Lab File: VX005862.D
Acq: 09 Nov 2018 15:32

Tgt Ion:152 Resp: 67136
Ion Ratio Lower Upper
152 100
115 56.0 39.4 118.1
150 157.1 0.0 346.4



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V





#74

N-amvl acetate

Concen: 1.412 ug/l

RT: 10.87 min Scan# 1594

Delta R.T. -0.02 min

Lab File: VX005862.D

Acq: 09 Nov 2018 15:32

Instrument :

MSVOA_X

ClientSampleId :

PB114659ZHE#05

Tot Ion: 43 Resp: 360





Ion Ratio Lower Upper

43 100

70 0.0 31.6 47.4#

55 0.0 19.7 29.5#

61 0.0 18.3 27.5#



#76

1,2,3-Trichloropropane

Concen: 23.612 ug/l

RT: 11.14 min Scan# 1638

Delta R.T. -0.16 min

Lab File: VX005862.D




Acq: 09 Nov 2018 15:32


Tgt Ion: 75 Resp: 32279

Ion Ratio Lower Upper

75 100

77 1.3 20.5 61.6#









#80
 1,3,5-Trimethylbenzene
 Concen: 1.774 ug/l
 RT: 11.51 min Scan# 1698
 Delta R.T. -0.00 min
 Lab File: VX005862.D
 Acq: 09 Nov 2018 15:32




Instrument :
 MSVOA_X
 ClientSampleId :
 PB114659ZHE#05


Tot Ion:105 Resp: 5854
 Ion Ratio Lower Upper
 105 100
 120 48.7 24.1 72.2



#81
 trans-1,4-Dichloro-2-butene
 Concen: 75.257 ug/l
 RT: 11.14 min Scan# 1638
 Delta R.T. 0.07 min
 Lab File: VX005862.D
 Acq: 09 Nov 2018 15:32

Tgt Ion: 75 Resp: 32279
 Ion Ratio Lower Upper
 75 100
 53 0.1 80.2 120.2#
 89 0.0 39.8 59.8#









#82
4-Chlorotoluene
Concen: 0.226 ug/l
RT: 11.51 min Scan# 1698
Delta R.T. -0.01 min
Lab File: VX005862.D
Acq: 09 Nov 2018 15:32




Instrument :
MSVOA_X
ClientSampleId :
PB114659ZHE#05


Tot Ion: 91 Resp: 737
Ion Ratio Lower Upper
91 100
126 0.0 15.2 45.6#



#83
tert-Butylbenzene
Concen: 0.238 ug/l
RT: 11.51 min Scan# 1698
Delta R.T. -0.26 min
Lab File: VX005862.D
Acq: 09 Nov 2018 15:32

Tgt Ion: 119 Resp: 793
Ion Ratio Lower Upper
119 100
91 92.9 28.5 85.5#
134 0.0 11.3 33.8#






#95
 Naphthalene
 Concen: 1.205 ug/l
 RT: 13.84 min Scan# 2081
 Delta R.T. 0.01 min
 Lab File: VX005862.D
 Acq: 09 Nov 2018 15:32

Instrument :
 MSVOA_X
 ClientSampleId :
 PB114659ZHE#05

Tot Ion	Ratio	Lower	Upper
128	100		
127	0.0	10.4	15.6#
129	0.0	8.7	13.1#



Abundance Ion 128.00 (127.70 to 128.70): V
 Ion 127.00 (126.70 to 127.70): V
 Ion 129.00 (128.70 to 129.70): V

