

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110918\
 Data File : VX005858.D
 Acq On : 09 Nov 2018 13:45
 Operator : JC/MD
 Sample : PB114659ZHE#01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB114659ZHE#01

Quant Time: Nov 12 00:22:33 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	114569	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	161924	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	148135	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	71034	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	71306	54.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.28%	
35) Dibromofluoromethane	5.50	113	56281	51.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.20%	
50) Toluene-d8	8.71	98	198388	54.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.36%	
62) 4-Bromofluorobenzene	11.14	95	67418	51.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.42%	
Target Compounds						
16) Acetone	2.44	43	7114	10.031	ug/l	94
20) Methylene Chloride	2.86	84	1565	1.343	ug/l	93
43) Isopropyl Acetate	6.46	43	27138	9.124	ug/l	99
80) 1,3,5-Trimethylbenzene	11.51	105	6330	1.813	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110918\
Data File : VX005858.D
Acq On : 09 Nov 2018 13:45
Operator : JC/MD
Sample : PB114659ZHE#01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB114659ZHE#01

Quant Time: Nov 12 00:22:33 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration

