

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110918\
 Data File : VX005859.D
 Acq On : 09 Nov 2018 14:12
 Operator : JC/MD
 Sample : PB114659ZHE#02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB114659ZHE#02

Quant Time: Nov 12 00:24:33 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	112421	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	158453	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	149062	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	70168	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	72065	55.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.52%	
35) Dibromofluoromethane	5.51	113	55297	51.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.62%	
50) Toluene-d8	8.71	98	198171	55.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.62%	
62) 4-Bromofluorobenzene	11.13	95	66771	51.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.66%	
Target Compounds						
16) Acetone	2.44	43	6182	8.884	ug/l	100
18) Methyl Acetate	2.78	43	5937	1.752	ug/l	99
20) Methylene Chloride	2.86	84	1409	1.233	ug/l	98
43) Isopropyl Acetate	6.47	43	26683	9.167	ug/l	98
80) 1,3,5-Trimethylbenzene	11.51	105	6153	1.784	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110918\
Data File : VX005859.D
Acq On : 09 Nov 2018 14:12
Operator : JC/MD
Sample : PB114659ZHE#02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB114659ZHE#02

Quant Time: Nov 12 00:24:33 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration

