

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110918\
 Data File : VX005861.D
 Acq On : 09 Nov 2018 15:05
 Operator : JC/MD
 Sample : PB114659ZHE#04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB114659ZHE#04

Quant Time: Nov 12 00:31:18 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	110857	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	161325	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	143166	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	67482	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	71315	55.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.92%	
35) Dibromofluoromethane	5.50	113	57456	52.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.72%	
50) Toluene-d8	8.71	98	191582	52.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.04%	
62) 4-Bromofluorobenzene	11.14	95	64231	48.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.94%	
Target Compounds						
16) Acetone	2.45	43	6365	9.276	ug/l	98
18) Methyl Acetate	2.78	43	10629	4.428	ug/l	100
20) Methylene Chloride	2.86	84	1456	1.292	ug/l	90
43) Isopropyl Acetate	6.46	43	25821	8.713	ug/l	98
80) 1,3,5-Trimethylbenzene	11.51	105	5971	1.801	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110918\
Data File : VX005861.D
Acq On : 09 Nov 2018 15:05
Operator : JC/MD
Sample : PB114659ZHE#04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB114659ZHE#04

Quant Time: Nov 12 00:31:18 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration

