

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110918\
 Data File : VX005863.D
 Acq On : 09 Nov 2018 15:58
 Operator : JC/MD
 Sample : PB114659ZHE#06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB114659ZHE#06

Quant Time: Nov 12 00:39:32 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	109512	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	148397	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	139297	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	64631	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	68540	54.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.88%	
35) Dibromofluoromethane	5.50	113	54259	53.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.52%	
50) Toluene-d8	8.71	98	185853	55.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.78%	
62) 4-Bromofluorobenzene	11.14	95	62352	51.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.36%	
Target Compounds						
16) Acetone	2.45	43	6189	9.130	ug/l	94
18) Methyl Acetate	2.78	43	12254	5.423	ug/l	97
20) Methylene Chloride	2.86	84	1483	1.332	ug/l	87
43) Isopropyl Acetate	6.45	43	26594	9.756	ug/l	99
80) 1,3,5-Trimethylbenzene	11.51	105	5449	1.716	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110918\
Data File : VX005863.D
Acq On : 09 Nov 2018 15:58
Operator : JC/MD
Sample : PB114659ZHE#06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB114659ZHE#06

Quant Time: Nov 12 00:39:32 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration

