

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110918\
 Data File : VX005866.D
 Acq On : 09 Nov 2018 17:18
 Operator : JC/MD
 Sample : J5876-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 48382-47191-47202

Quant Time: Nov 12 00:46:01 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	109800	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	152286	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	141883	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	66125	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	69726	55.24	ug/l	0.00
Spiked Amount			Recovery	=	110.48%	
35) Dibromofluoromethane	5.50	113	56313	54.37	ug/l	0.00
Spiked Amount			Recovery	=	108.74%	
50) Toluene-d8	8.71	98	191397	55.58	ug/l	0.00
Spiked Amount			Recovery	=	111.16%	
62) 4-Bromofluorobenzene	11.14	95	62168	50.21	ug/l	0.00
Spiked Amount			Recovery	=	100.42%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.44	43	9738	14.328	ug/l	92
18) Methyl Acetate	2.78	43	10456	4.388	ug/l	98
20) Methylene Chloride	2.86	84	3458	3.097	ug/l #	77
43) Isopropyl Acetate	6.47	43	19902	7.114	ug/l	99
95) Naphthalene	13.83	128	3080	1.810	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110918\
Data File : VX005866.D
Acq On : 09 Nov 2018 17:18
Operator : JC/MD
Sample : J5876-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
48382-47191-47202

Quant Time: Nov 12 00:46:01 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration

