

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110918\
 Data File : VX005868.D
 Acq On : 09 Nov 2018 18:11
 Operator : JC/MD
 Sample : J5884-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-8A-S

Quant Time: Nov 12 00:52:27 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	107708	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	145548	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	139739	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	65382	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	69460	56.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.20%	
35) Dibromofluoromethane	5.51	113	56842	57.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.84%	
50) Toluene-d8	8.71	98	187857	57.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.16%	
62) 4-Bromofluorobenzene	11.14	95	61478	51.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.92%	
Target Compounds						
16) Acetone	2.45	43	11254	16.880	ug/l	99
18) Methyl Acetate	2.78	43	5963	1.911	ug/l	97
20) Methylene Chloride	2.86	84	22883	20.895	ug/l	94
25) 2-Butanone	4.70	43	2661	2.539	ug/l #	87
43) Isopropyl Acetate	6.46	43	15364	5.747	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110918\
Data File : VX005868.D
Acq On : 09 Nov 2018 18:11
Operator : JC/MD
Sample : J5884-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-8A-S

Quant Time: Nov 12 00:52:27 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration

