

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110918\
 Data File : VX005875.D
 Acq On : 09 Nov 2018 22:11
 Operator : JC/MD
 Sample : VX1109MBL02
 Misc : 5.00µ/10mL/100µL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1109MBL02

Quant Time: Nov 12 01:19:38 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	106271	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	6.86	114	151448	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	124149	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	60199	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	65427	53.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.10%	
35) Dibromofluoromethane	5.51	113	52302	50.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.54%	
50) Toluene-d8	8.72	98	165589	48.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.70%	
62) 4-Bromofluorobenzene	11.14	95	54556	44.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.62%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110918\
Data File : VX005875.D
Acq On : 09 Nov 2018 22:11
Operator : JC/MD
Sample : VX1109MBL02
Misc : 5.00µL/10mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1109MBL02

Quant Time: Nov 12 01:19:38 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration

