

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110918\
 Data File : VX005886.D
 Acq On : 10 Nov 2018 03:04
 Operator : JC/MD
 Sample : J5889-12
 Misc : 20.29g/5mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BOTTOM-C

Quant Time: Nov 12 09:08:00 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	133136	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	182921	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	164990	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	90396	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	74495	48.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.34%	
35) Dibromofluoromethane	5.50	113	60411	48.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.12%	
50) Toluene-d8	8.71	98	216081	52.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.48%	
62) 4-Bromofluorobenzene	11.14	95	80431	54.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.16%	
Target Compounds						
84) 1,2,4-Trimethylbenzene	11.81	105	21293	4.234	ug/l	99
89) n-Butylbenzene	12.39	91	5147	1.141	ug/l #	50
95) Naphthalene	13.83	128	278592	43.006	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110918\
Data File : VX005886.D
Acq On : 10 Nov 2018 03:04
Operator : JC/MD
Sample : J5889-12
Misc : 20.29g/5mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BOTTOM-C

Quant Time: Nov 12 09:08:00 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration

