

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110918\
 Data File : VX005888.D
 Acq On : 10 Nov 2018 03:58
 Operator : JC/MD
 Sample : J5889-14
 Misc : 13.32g/5mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BOTTOM-E

Quant Time: Nov 12 09:09:46 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	134010	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	200935	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	166567	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	88090	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	75778	49.19	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.38%	
35) Dibromofluoromethane	5.50	113	62948	46.06	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.12%	
50) Toluene-d8	8.71	98	219890	48.40	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.80%	
62) 4-Bromofluorobenzene	11.14	95	79402	48.61	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.22%	
Target Compounds						
84) 1,2,4-Trimethylbenzene	11.81	105	16434	3.328	ug/l	98
95) Naphthalene	13.83	128	205381	32.821	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110918\
Data File : VX005888.D
Acq On : 10 Nov 2018 03:58
Operator : JC/MD
Sample : J5889-14
Misc : 13.32g/5mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BOTTOM-E

Quant Time: Nov 12 09:09:46 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration

