

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX110918\
 Data File : VX005872.D
 Acq On : 09 Nov 2018 19:58
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 12 01:14:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	91	0.00
2 T	Dichlorodifluoromethane	0.451	0.467	-3.5	92	0.00
3 P	Chloromethane	0.592	0.642	-8.4	101	0.00
4 C	Vinyl Chloride	0.594	0.599	-0.8#	88	0.00
5 T	Bromomethane	0.363	0.310	14.6	78	0.00
6 T	Chloroethane	0.354	0.354	0.0	91	0.00
7 T	Trichlorofluoromethane	0.788	0.796	-1.0	90	0.00
8 T	Diethyl Ether	0.309	0.299	3.2	91	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.453	0.443	2.2	90	0.00
10 T	Methyl Iodide	0.550	0.550	0.0	84	0.00
11 T	Tert butyl alcohol	0.165	0.162	1.8	92	0.00
12 CM	1,1-Dichloroethene	0.442	0.422	4.5#	88	0.00
13 T	Acrolein	0.070	0.074	-5.7	89	0.00
14 T	Allyl chloride	0.987	0.979	0.8	89	0.00
15 T	Acrylonitrile	0.341	0.354	-3.8	91	0.00
16 T	Acetone	0.309	0.322	-4.2	93	0.00
17 T	Carbon Disulfide	1.318	1.267	3.9	85	0.00
18 T	Methyl Acetate	0.921	0.982	-6.6	104	0.00
19 T	Methyl tert-butyl Ether	1.508	1.590	-5.4	94	0.00
20 T	Methylene Chloride	0.508	0.478	5.9	90	0.00
21 T	trans-1,2-Dichloroethene	0.466	0.461	1.1	88	0.00
22 T	Diisopropyl ether	1.729	1.780	-2.9	93	0.00
23 T	Vinyl Acetate	1.483	1.604	-8.2	96	-0.01
24 P	1,1-Dichloroethane	0.922	0.939	-1.8	94	0.00
25 T	2-Butanone	0.487	0.487	0.0	87	0.00
26 T	2,2-Dichloropropane	0.677	0.635	6.2	83	0.00
27 T	cis-1,2-Dichloroethene	0.515	0.514	0.2	88	0.00
28 T	Bromochloromethane	0.412	0.433	-5.1	85	0.00
29 T	Tetrahydrofuran	0.312	0.337	-8.0	91	0.00
30 C	Chloroform	0.890	0.887	0.3#	89	0.00
31 T	Cyclohexane	0.801	0.812	-1.4	88	0.00
32 T	1,1,1-Trichloroethane	0.742	0.778	-4.9	91	0.00
33 S	1,2-Dichloroethane-d4	0.575	0.576	-0.2	90	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	90	0.00
35 S	Dibromofluoromethane	0.340	0.345	-1.5	87	0.00
36 T	1,1-Dichloropropene	0.482	0.497	-3.1	89	0.00
37 T	Ethyl Acetate	0.634	0.662	-4.4	90	0.00
38 T	Carbon Tetrachloride	0.483	0.517	-7.0	94	0.00
39 T	Methylcyclohexane	0.489	0.523	-7.0	90	0.00
40 TM	Benzene	1.405	1.554	-10.6	96	0.00
41 T	Methacrylonitrile	0.343	0.362	-5.5	91	0.00
42 TM	1,2-Dichloroethane	0.503	0.538	-7.0	94	0.00
43 T	Isopropyl Acetate	0.918	1.013	-10.3	93	0.00
44 TM	Trichloroethene	0.357	0.372	-4.2	93	0.00
45 C	1,2-Dichloropropane	0.341	0.364	-6.7#	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX110918\
 Data File : VX005872.D
 Acq On : 09 Nov 2018 19:58
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 12 01:14:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.214	0.224	-4.7	94	0.00
47 T	Bromodichloromethane	0.404	0.445	-10.1	97	0.00
48 T	Methyl methacrylate	0.412	0.456	-10.7	94	0.00
49 T	1,4-Dioxane	0.009	0.010	-11.1	99	0.00
50 S	Toluene-d8	1.131	1.161	-2.7	91	0.00
51 T	4-Methyl-2-Pentanone	0.579	0.621	-7.3	96	0.00
52 CM	Toluene	0.785	0.803	-2.3#	91	0.00
53 T	t-1,3-Dichloropropene	0.477	0.492	-3.1	83	0.00
54 T	cis-1,3-Dichloropropene	0.486	0.535	-10.1	96	0.00
55 T	1,1,2-Trichloroethane	0.350	0.333	4.9	89	0.00
56 T	Ethyl methacrylate	0.524	0.526	-0.4	85	0.00
57 T	1,3-Dichloropropane	0.588	0.557	5.3	83	0.00
58 T	2-Chloroethyl Vinyl ether	0.257	0.315	-22.6#	101	0.00
59 T	2-Hexanone	0.510	0.495	2.9	81	0.00
60 T	Dibromochloromethane	0.378	0.357	5.6	79	0.00
61 T	1,2-Dibromoethane	0.387	0.352	9.0	80	0.00
62 S	4-Bromofluorobenzene	0.406	0.436	-7.4	93	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	84	0.00
64 T	Tetrachloroethene	0.426	0.400	6.1	72	0.00
65 PM	Chlorobenzene	1.002	1.024	-2.2	82	0.00
66 T	1,1,1,2-Tetrachloroethane	0.348	0.378	-8.6	91	0.00
67 C	Ethyl Benzene	1.606	1.776	-10.6#	86	0.00
68 T	m/p-Xylenes	0.613	0.686	-11.9	93	0.00
69 T	o-Xylene	0.549	0.657	-19.7	96	0.00
70 T	Styrene	0.906	1.106	-22.1#	96	0.00
71 P	Bromoform	0.292	0.342	-17.1	97	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
73 T	Isopropylbenzene	2.891	3.192	-10.4	96	0.00
74 T	N-amyl acetate	1.473	1.631	-10.7	98	0.00
75 P	1,1,2,2-Tetrachloroethane	1.061	1.105	-4.1	97	0.00
76 T	1,2,3-Trichloropropane	1.018	1.066	-4.7	95	0.00
77 T	Bromobenzene	0.815	0.836	-2.6	95	0.00
78 T	n-propylbenzene	3.408	3.703	-8.7	95	0.00
79 T	2-Chlorotoluene	2.060	2.187	-6.2	95	0.00
80 T	1,3,5-Trimethylbenzene	2.457	2.755	-12.1	96	0.00
81 T	trans-1,4-Dichloro-2-butene	0.319	0.343	-7.5	97	0.00
82 T	4-Chlorotoluene	2.427	2.619	-7.9	96	0.00
83 T	tert-Butylbenzene	2.486	2.697	-8.5	98	0.00
84 T	1,2,4-Trimethylbenzene	2.543	2.812	-10.6	95	0.00
85 T	sec-Butylbenzene	3.040	3.269	-7.5	96	0.00
86 T	p-Isopropyltoluene	2.693	2.930	-8.8	95	0.00
87 T	1,3-Dichlorobenzene	1.564	1.553	0.7	95	0.00
88 T	1,4-Dichlorobenzene	1.601	1.575	1.6	95	0.00
89 T	n-Butylbenzene	2.495	2.591	-3.8	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX110918\
Data File : VX005872.D
Acq On : 09 Nov 2018 19:58
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Nov 12 01:14:29 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.450	0.495	-10.0	100	0.00
91 T	1,2-Dichlorobenzene	1.567	1.576	-0.6	96	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.259	0.284	-9.7	96	0.00
93 T	1,2,4-Trichlorobenzene	1.103	1.175	-6.5	94	0.00
94 T	Hexachlorobutadiene	0.579	0.581	-0.3	93	0.00
95 T	Naphthalene	3.195	3.675	-15.0	96	0.00
96 T	1,2,3-Trichlorobenzene	1.124	1.209	-7.6	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6