

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110920\
 Data File : VX019366.D
 Acq On : 09 Nov 2020 14:36
 Operator : JC/MD
 Sample : L4670-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 B-17-MW-4

Quant Time: Nov 10 02:15:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 27 13:55:07 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	303501	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	537240	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	513789	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	261626	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	214936	54.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.82%	
35) Dibromofluoromethane	5.46	113	167996	49.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.56%	
50) Toluene-d8	8.70	98	668520	52.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.10%	
62) 4-Bromofluorobenzene	11.12	95	260162	54.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.00%	
Target Compounds						
3) Chloromethane	1.31	50	3760	1.330	ug/l	98
16) Acetone	2.42	43	6304	4.669	ug/l	96
20) Methylene Chloride	2.84	84	1598	0.471	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110920\
Data File : VX019366.D
Acq On : 09 Nov 2020 14:36
Operator : JC/MD
Sample : L4670-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
B-17-MW-4

Quant Time: Nov 10 02:15:18 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 13:55:07 2020
Response via : Initial Calibration

