

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110920\  
 Data File : VX019369.D  
 Acq On : 09 Nov 2020 15:44  
 Operator : JC/MD  
 Sample : L4668-03  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 14 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 MW-3

Quant Time: Nov 10 02:44:39 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X102720W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Oct 27 13:55:07 2020  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc  | Units   | Dev(Min) |
|-----------------------------|-------|------|----------|-------|---------|----------|
| 1) Pentafluorobenzene       | 5.64  | 168  | 288494   | 50.00 | ug/l    | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.83  | 114  | 500067   | 50.00 | ug/l    | 0.00     |
| 63) Chlorobenzene-d5        | 10.10 | 117  | 481049   | 50.00 | ug/l    | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.06 | 152  | 247516   | 50.00 | ug/l    | 0.00     |
| System Monitoring Compounds |       |      |          |       |         |          |
| 33) 1,2-Dichloroethane-d4   | 6.03  | 65   | 202498   | 54.43 | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =     | 108.86% |          |
| 35) Dibromofluoromethane    | 5.46  | 113  | 156519   | 49.83 | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =     | 99.66%  |          |
| 50) Toluene-d8              | 8.70  | 98   | 622005   | 52.53 | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =     | 105.06% |          |
| 62) 4-Bromofluorobenzene    | 11.12 | 95   | 243692   | 54.34 | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =     | 108.68% |          |
| Target Compounds            |       |      |          |       |         |          |
| 3) Chloromethane            | 1.31  | 50   | 2779     | 1.034 | ug/l    | 96       |
| 16) Acetone                 | 2.42  | 43   | 6238     | 4.860 | ug/l    | 98       |
| 20) Methylene Chloride      | 2.84  | 84   | 1564     | 0.485 | ug/l    | 91       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110920\  
Data File : VX019369.D  
Acq On : 09 Nov 2020 15:44  
Operator : JC/MD  
Sample : L4668-03  
Misc : 5.0mL/MSVOA X/WATER  
ALS Vial : 14 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
MW-3

Quant Time: Nov 10 02:44:39 2020  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_X\METHOD\82X102720W.M  
Quant Title : SW846 8260  
QLast Update : Tue Oct 27 13:55:07 2020  
Response via : Initial Calibration

