


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX110920\
 Data File : VX019356.D
 Acq On : 09 Nov 2020 08:50
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
 Title : SW846 8260
 Last Update : Tue Oct 27 13:55:07 2020



AutoFind: Scans 1645, 1646, 1647; Background Corrected with Scan 1638

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.4	9624	PASS
75	95	30	60	49.8	27576	PASS
95	95	100	100	100.0	55362	PASS
96	95	5	9	6.8	3760	PASS
173	174	0.00	2	0.9	379	PASS
174	95	50	100	76.9	42546	PASS
175	174	5	9	7.2	3077	PASS
176	174	95	101	98.1	41757	PASS
177	176	5	9	6.5	2694	PASS