

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110921\
 Data File : VX025122.D
 Acq On : 09 Nov 2021 15:20
 Operator : JC/MD
 Sample : PB140605TB 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VLEB605

Quant Time: Nov 10 02:53:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML110821WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Nov 10 02:50:07 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	135012	50.000	ug/L	# 0.00
28) Chlorobenzene-d5	10.055	117	122750	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	48065	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.367	65	51754	42.393	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	84.780%	
7) Chloroethane-d5	1.666	69	44433	58.056	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	116.120%	
11) 1,1-Dichloroethene-d2	2.306	63	81871	34.817	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	69.640%	
21) 2-Butanone-d5	4.464	46	84815	101.088	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	101.090%	
24) Chloroform-d	5.068	84	107612	44.837	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	89.680%	
26) 1,2-Dichloroethane-d4	5.964	65	79298	50.505	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	101.020%	
32) Benzene-d6	5.982	84	179376	41.042	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	82.080%	
36) 1,2-Dichloropropane-d6	7.312	67	63489	49.299	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	98.600%	
41) Toluene-d8	8.653	98	161366	43.921	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	87.840%	
43) trans-1,3-Dichloroprop...	8.951	79	31420	45.712	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	91.420%	
47) 2-Hexanone-d5	9.390	63	57090	103.780	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	103.780%	
56) 1,1,2,2-Tetrachloroeth...	11.195	84	80258	45.022	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	90.040%	
66) 1,2-Dichlorobenzene-d4	12.323	152	45857	48.675	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	97.340%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110921\
Data File : VX025122.D
Acq On : 09 Nov 2021 15:20
Operator : JC/MD
Sample : PB140605TB 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VLEB605

Quant Time: Nov 10 02:53:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML110821WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Nov 10 02:50:07 2021
Response via : Initial Calibration

