

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110922\
 Data File : VX032676.D
 Acq On : 09 Nov 2022 13:58
 Operator : JC/MD
 Sample : N5515-01 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FMI121

Quant Time: Nov 10 03:04:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	113825	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	188354	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	169194	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	82214	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	79853	50.490	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 100.980%			
35) Dibromofluoromethane	5.385	113	61250	45.588	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 91.180%			
50) Toluene-d8	8.647	98	228822	48.979	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 97.960%			
62) 4-Bromofluorobenzene	11.079	95	90731	49.718	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 99.440%			
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.794	84	934	0.709	ug/l #	82
39) Methylcyclohexane	7.379	83	2050	0.926	ug/l #	87
52) Toluene	8.714	92	4128	1.195	ug/l	95
67) Ethyl Benzene	10.195	91	46316	6.834	ug/l	97
68) m/p-Xylenes	10.299	106	71423	27.139	ug/l	99
69) o-Xylene	10.640	106	23678	9.254	ug/l	95
73) Isopropylbenzene	10.963	105	9605	1.474	ug/l	99
78) n-propylbenzene	11.305	91	25049	3.337	ug/l	98
80) 1,3,5-Trimethylbenzene	11.451	105	34243	6.150	ug/l	97
84) 1,2,4-Trimethylbenzene	11.750	105	126993	22.932	ug/l	99
89) n-Butylbenzene	12.329	91	8155	1.652	ug/l #	42
95) Naphthalene	13.774	128	41306	6.400	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110922\
Data File : VX032676.D
Acq On : 09 Nov 2022 13:58
Operator : JC/MD
Sample : N5515-01 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FMI121

Quant Time: Nov 10 03:04:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

