

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110922\
 Data File : VX032677.D
 Acq On : 09 Nov 2022 14:21
 Operator : JC/MD
 Sample : N5514-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-10

Quant Time: Nov 10 03:04:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	120629	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	200743	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	177430	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	82544	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	83663	49.915	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.820%
35) Dibromofluoromethane	5.391	113	64948	45.357	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.720%
50) Toluene-d8	8.647	98	243279	48.860	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.720%
62) 4-Bromofluorobenzene	11.079	95	91180	46.880	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110922\
Data File : VX032677.D
Acq On : 09 Nov 2022 14:21
Operator : JC/MD
Sample : N5514-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-10

Quant Time: Nov 10 03:04:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

