

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110922\
 Data File : VX032680.D
 Acq On : 09 Nov 2022 15:31
 Operator : JC/MD
 Sample : N5551-13 40X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-8I

Quant Time: Nov 10 03:05:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	112973	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	189116	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	170453	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83180	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	78841	50.226	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.460%
35) Dibromofluoromethane	5.385	113	61546	45.624	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.240%
50) Toluene-d8	8.647	98	229914	49.015	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.020%
62) 4-Bromofluorobenzene	11.079	95	89339	48.758	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.520%
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.794	84	712	0.544	ug/l #	76
31) Cyclohexane	5.477	56	7918	4.160	ug/l	89
39) Methylcyclohexane	7.391	83	3613	1.626	ug/l	90
40) Benzene	6.037	78	38310	7.207	ug/l	98
52) Toluene	8.720	92	26523	7.647	ug/l	99
67) Ethyl Benzene	10.195	91	311254	45.586	ug/l	100
68) m/p-Xylenes	10.299	106	70431	26.565	ug/l	100
69) o-Xylene	10.640	106	15750	6.110	ug/l	100
73) Isopropylbenzene	10.963	105	14292	2.168	ug/l	98
78) n-propylbenzene	11.305	91	44598	5.873	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	7185	1.275	ug/l	97
84) 1,2,4-Trimethylbenzene	11.750	105	33663	6.008	ug/l	99
89) n-Butylbenzene	12.335	91	2116	0.424	ug/l #	56
95) Naphthalene	13.774	128	79408	12.160	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110922\
Data File : VX032680.D
Acq On : 09 Nov 2022 15:31
Operator : JC/MD
Sample : N5551-13 40X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-8I

Quant Time: Nov 10 03:05:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

