

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110922\
 Data File : VX032684.D
 Acq On : 09 Nov 2022 17:01
 Operator : JC/MD
 Sample : N5512-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 AUD-1911-TOTE

Quant Time: Nov 10 03:05:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	106958	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	178332	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	154553	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	73633	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	76229	51.293	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.580%
35) Dibromofluoromethane	5.385	113	59177	46.521	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.040%
50) Toluene-d8	8.647	98	213474	48.262	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.520%
62) 4-Bromofluorobenzene	11.079	95	83485	48.318	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.640%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	4467	8.270	ug/l	95
68) m/p-Xylenes	10.299	106	1290	0.537	ug/l	94
84) 1,2,4-Trimethylbenzene	11.750	105	3514	0.708	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110922\
Data File : VX032684.D
Acq On : 09 Nov 2022 17:01
Operator : JC/MD
Sample : N5512-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
AUD-1911-TOTE

Quant Time: Nov 10 03:05:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

