

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110922\
 Data File : VX032685.D
 Acq On : 09 Nov 2022 17:24
 Operator : JC/MD
 Sample : N5512-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 AUD-1896-TOTE

Quant Time: Nov 10 03:06:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	118658	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	196108	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	177731	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	84791	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	82629	50.117	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.240%
35) Dibromofluoromethane	5.391	113	63979	45.737	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.480%
50) Toluene-d8	8.647	98	238552	49.043	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.080%
62) 4-Bromofluorobenzene	11.079	95	93563	49.242	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.480%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	2351	3.923	ug/l #	85
19) Methyl tert-butyl Ether	3.117	73	3038	0.726	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110922\
Data File : VX032685.D
Acq On : 09 Nov 2022 17:24
Operator : JC/MD
Sample : N5512-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
AUD-1896-TOTE

Quant Time: Nov 10 03:06:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

