

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110922\
 Data File : VX032688.D
 Acq On : 09 Nov 2022 18:34
 Operator : JC/MD
 Sample : N5551-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-8

Quant Time: Nov 10 03:06:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	109608	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	182440	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	156855	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	71886	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	77528	50.906	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.820%
35) Dibromofluoromethane	5.385	113	59824	45.970	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.940%
50) Toluene-d8	8.653	98	219981	48.613	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.220%
62) 4-Bromofluorobenzene	11.079	95	81062	45.859	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.720%
Target Compounds						
					Qvalue	
67) Ethyl Benzene	10.195	91	6560	1.044	ug/l	96
68) m/p-Xylenes	10.299	106	2798	1.147	ug/l	95
69) o-Xylene	10.646	106	562	0.237	ug/l	92
84) 1,2,4-Trimethylbenzene	11.756	105	2760	0.570	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110922\
Data File : VX032688.D
Acq On : 09 Nov 2022 18:34
Operator : JC/MD
Sample : N5551-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-8

Quant Time: Nov 10 03:06:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

