

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110922\
 Data File : VX032704.D
 Acq On : 10 Nov 2022 04:08
 Operator : JC/MD
 Sample : N5502-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 2182

Quant Time: Nov 10 05:21:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	128004	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	212301	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	198065	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	93342	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	91507	51.449	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.900%
35) Dibromofluoromethane	5.385	113	70067	46.268	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.540%
50) Toluene-d8	8.647	98	265572	50.434	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.860%
62) 4-Bromofluorobenzene	11.079	95	105093	51.092	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.180%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	8185	12.662	ug/l	97
20) Methylene Chloride	2.788	84	6247	4.215	ug/l	88
37) Ethyl Acetate	4.714	43	14616	6.659	ug/l	99
43) Isopropyl Acetate	6.336	43	70860	21.024	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110922\
Data File : VX032704.D
Acq On : 10 Nov 2022 04:08
Operator : JC/MD
Sample : N5502-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
2182

Quant Time: Nov 10 05:21:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

