

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110922\
 Data File : VX032707.D
 Acq On : 10 Nov 2022 05:18
 Operator : JC/MD
 Sample : N5506-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-8

Quant Time: Nov 10 05:44:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	129434	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	217654	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	205182	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	94666	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	90332	50.228	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.460%	
35) Dibromofluoromethane	5.385	113	70846	45.632	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 91.260%	
50) Toluene-d8	8.647	98	271742	50.336	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.680%	
62) 4-Bromofluorobenzene	11.079	95	106456	50.482	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.960%	
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	8468	12.955	ug/l	100
20) Methylene Chloride	2.788	84	5391	3.597	ug/l	90
37) Ethyl Acetate	4.715	43	15470	6.874	ug/l	98
43) Isopropyl Acetate	6.336	43	71062	20.566	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110922\
Data File : VX032707.D
Acq On : 10 Nov 2022 05:18
Operator : JC/MD
Sample : N5506-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-8

Quant Time: Nov 10 05:44:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

