

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110922\
 Data File : VX032708.D
 Acq On : 10 Nov 2022 05:41
 Operator : JC/MD
 Sample : N5506-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-9

Quant Time: Nov 10 08:08:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	122033	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	204719	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	197445	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95437	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	86493	51.010	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.020%
35) Dibromofluoromethane	5.385	113	66983	45.870	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.740%
50) Toluene-d8	8.647	98	257726	50.756	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.520%
62) 4-Bromofluorobenzene	11.079	95	103906	52.386	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.780%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	10776	17.486	ug/l	97
20) Methylene Chloride	2.788	84	4799	3.396	ug/l	92
37) Ethyl Acetate	4.714	43	15173	7.168	ug/l	99
43) Isopropyl Acetate	6.342	43	70466	21.682	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110922\
Data File : VX032708.D
Acq On : 10 Nov 2022 05:41
Operator : JC/MD
Sample : N5506-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-9

Quant Time: Nov 10 08:08:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

