

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110922\
 Data File : VX032710.D
 Acq On : 10 Nov 2022 06:27
 Operator : JC/MD
 Sample : N5509-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-7

Quant Time: Nov 10 08:08:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	123441	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	203858	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	191328	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90534	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	86792	50.602	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.200%
35) Dibromofluoromethane	5.385	113	68004	46.766	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.540%
50) Toluene-d8	8.647	98	255703	50.570	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.140%
62) 4-Bromofluorobenzene	11.079	95	100949	51.110	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.220%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	10875	17.445	ug/l	100
20) Methylene Chloride	2.794	84	6598	4.616	ug/l #	95
37) Ethyl Acetate	4.715	43	15896	7.542	ug/l	98
43) Isopropyl Acetate	6.342	43	72992	22.554	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110922\
Data File : VX032710.D
Acq On : 10 Nov 2022 06:27
Operator : JC/MD
Sample : N5509-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-7

Quant Time: Nov 10 08:08:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

