

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110922\
 Data File : VX032711.D
 Acq On : 10 Nov 2022 06:50
 Operator : JC/MD
 Sample : N5509-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-13

Quant Time: Nov 10 08:09:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	124081	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	209020	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	193404	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92840	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	88912	51.571	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.140%
35) Dibromofluoromethane	5.385	113	67518	45.285	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.560%
50) Toluene-d8	8.647	98	259902	50.132	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.260%
62) 4-Bromofluorobenzene	11.079	95	103340	51.028	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.060%
Target Compounds						
16) Acetone	2.386	43	12668	20.216	ug/l	98
20) Methylene Chloride	2.794	84	6774	4.715	ug/l	94
37) Ethyl Acetate	4.714	43	15565	7.202	ug/l	97
43) Isopropyl Acetate	6.342	43	72850	21.954	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110922\
Data File : VX032711.D
Acq On : 10 Nov 2022 06:50
Operator : JC/MD
Sample : N5509-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-13

Quant Time: Nov 10 08:09:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

